



Formaldehyde on Glass Fibre Filters

Frequency Count of Results for this Analyte



Results for this Laboratory



Summary of your results and their ratio to the mean

sample number	your result	'true' result	Difference(%)	standardised results	RPI	performance category
1	4.40	4.73	-7.02	0.930	< 79	1
2	6.46	6.51	-0.81	0.992	79 to 330	2
3	19.44	19.65	-1.08	0.989	> 330	3
4	13.75	13.76	-0.10	0.999		

Reference Value Used: No mean: 0.98

RPI Reference Value: 183

Results for Your Laboratory:

Number of outliers: 0
 Performance Index (PI): 13
 Running Performance Index (RPI): 5
 Performance Category: 1

Results for All Participants:

Mean PI: 69.97 Median PI: 12
 Mean RPI: 39.79 Median RPI: 13

Your laboratory is ranked 4 out of 47 labs.

additional statistical data (see the statistical protocol paragraphs 12 to 18)

analysis of variance: Delta: -0.022528 SSW: 0.003077 Coefficient of variation: 3.090

RPI Plot for your Laboratory

